

EMLMF inter-comparison program (2/2010)

for the calibration of
Burette (25 mL)
Using gravimetric method

Summary of the results

Laboratory No.	Calculated volume at 20 °C (ml)	Uncertainty (± ml)
1	25.15	0.10
2	25.15	0.10
3	25.10	0.06
4	25.16	0.14
5	25.18	0.09
6	25.15	1.05

**Results of EMLMF Intercomparison Program No. (2/2010)
Burrte (25 mL)**

