

EMLMF inter-comparison program (3/2010)

for the calibration of
Graduated Cylinder (5 mL)
Using gravimetric method

Summary of the results

Laboratory No.	Calculated volume at 20 °C (ml)	Uncertainty (± ml)
1	5.02	0.100
2	5.02	0.110
3	5.007	0.067
4	5.02	0.130
5	5.02	0.059
6	5.02	0.066

**Results of EMLMF Intercomparison Program No. (3/2010)
Graduated Cylinder (5 mL)**

